

# **ASOPS ENGINE DATA SHEET**

NEXT GENERATION, TIME-DOMAIN

SPECTROSCOPY SYSTEM

#### PRECISE AND FAST

As the leading supplier for ASynchronous OPtical Sampling (ASOPS) technology for almost 15 years, Laser Quantum's third generation ASOPS technique allows for the most precise and fastest time-domain spectroscopy available on the market today. The ASOPS Engine includes all components necessary for high speed ASOPS time domain spectroscopy: two femtosecond lasers (sold separately), master and slave, in addition to the TL-1000 ASOPS electronic unit for stabilizing the repetition rate of the slave laser with an offset to its master laser, and an optical trigger generation unit.




Fig. 1. Two-dimensional thickness measurement of a Si/Mo superlattice structure using ASOPS. At each pixel the time-resolved reflectivity change is measured and the superlattice period can be directly extracted from photoinduced ultrasonics with sub-nm resolution. The fast acquisition time allows quick measurement of a two dimensional 100x100 pixel scan with two adjacent pixels separated by 100 µm.



Turn-key Ti:Sapphire 1 GHz lasers

Key features of the ASOPS Engine include extremely fast acquisition speed of up to 20 kHz, an unprecedented time-resolution of significantly below 60 fs\* for 1 GHz lasers and below 100 fs for 84 MHz lasers respectively. This unique combination of high scan rates, long measurement windows and excellent time resolution is impossible with conventional time-domain spectrometers.

 $^{\ast}$  Typical values for the time resolution are <45 fs @ 1 GHz and < 100 fs at 84 MHz.




Fig. 2. Fourier transform of reflectivity measurements taken on ZnO samples with an ASOPS system. The excellent time stability provided by the TL-1000 ASOPS control unit allows it to measure even high phonon frequencies such as the observed 13.15 THz phonon mode in ZnO.

#### **ASOPS ENGINE LASER DATASHEET**

## **ASOPS TECHNOLOGY**

The key feature of the ASOPS technology is the use of two femtosecond lasers with repetition rates f<sub>R</sub> locked together in a master-slave configuration with a slight offset  $\Delta f_R$ . This offset, typically between 10 Hz and 10 kHz, causes the delay among pairs of pulses from the lasers to incrementally increase by  $\Delta \tau = \Delta f_R / f_R^2$  with each shot, for example, a 10 fs increase at  $\Delta f_R = 10$  kHz. If the lasers are then used as pump and probe lasers timedelay happens automatically, and the delay  $\tau$  between pump and probe pulse pairs undergoes a linear ramp  $\tau = txf_R/\Delta f_R$  as function of real time t, replicating itself at a rate given by  $\Delta f_R$ . Figure 3 illustrates the principle for an optical pump-probe time domain spectrometer (TDS) setup. The lasers are then used as they would be in a classical setup except that no translation stage is required. Timing precision is now determined by the ability to measure and stabilise the repetition rate offset. Uncertainties at the level of a few parts in 10<sup>5</sup> are reached — typically more than an order of magnitude better than mechanical delay generators.




Fig. 3. Realisation of the time delay increment of an ASOPS system. The pulse to pulse separation for pump and probe pulse trains differs by the time increment which increases linearly over time as seen in the above graph. Very short scan times of the order 100 µs can be realised with a measurement window of 1 ns using the ASOPS technique based on our 1 GHz taccor laser range. If 84 MHz lasers are employed, much longer scan delays of up to 12 ns can be realised.

## **EXCELLENT TIME-RESOLUTION**

Excellent time-resolution of significantly below 60fs for 1 GHz lasers and significantly below 100 fs for 84 MHz lasers is ensured by stabilising the master and slave lasers using the third generation TL-1000 ASOPS unit based on the patented DDS technology. The ASOPS Engine is available with the range of 1 GHz taccor lasers as well as the 84 MHz lasers from our venteon and gecco series. Typical time resolution values for both MHz and GHz ASOPS Engine systems can be seen in Figure 4 and 5 respectively.




Fig. 4. Measured timing jitter for a 1 GHz based ASOPS system for offset frequencies between 2 and 5 kHz showing 60 fs or better. At larger offset frequencies the time resolution increases due to the limited bandwidth of the data acquisition card.




Fig. 5. Measured timing jitter for a 84 MHz based ASOPS system and offset frequencies between 90 Hz and 500 Hz. Note that at larger offset frequencies the time resolution is worse as the required real time bandwidth becomes comparable to the repetition rate.

#### **ASOPS ENGINE LASER DATASHEET**

### **APPLICATIONS**

#### **High-resolution terahertz spectroscopy**

Spectroscopy in the 0.1-10 THz range of the electromagnetic spectrum is an important application that relies on ultrafast optical TDS, where precision and speed are key factors in many applications. Terahertz spectroscopy has great potential for use in gas spectroscopy and sensing, explosives and drug detection and monitoring, in-line paper or foil thickness measurements, solar-cell inspection and many more. Employing the ASOPS technology can be used to perform measurements with high frequency precision at a resolution of 1 GHz. An absorption spectrum of water vapour in the THz frequency range is shown in Figure 6.

# Ultrafast THz-spectroscopy in transient multitesla magnetic fields

The rapid data acquisition capability (up to 10,000 single-scan traces per second at 1 GHz resolution) permits the investigation of dynamic processes and studies under rapidly varying environmental conditions. An experiment was performed using the ASOPS Engine for THz spectroscopy where 100 THz traces were recorded while a magnetic field pulse of several tesla was applied to the sample. In this experiment the cyclotron resonance of a 2D electron gas in millisecond GaAs/AlGaAs heterostructure could be recorded as the magnetic field evolved on the ms timescale.

Figure 7 shows the evolution of the sample transmission and the measured magnetic field as function of time with a clear signature of the cyclotron resonance and correlation with the magnetic field. Electron mobility and effective values can directly be calculated from this measurement.

# Spatially resolved wafer mapping via picosecond ultrasound

A common method for wafer metrology or monitoring the growth of multilayer nanostructures is the use of laser-induced picosecond ultrasound, where a strong laser pulse launches a heat wave (very high-frequency ultrasound) into a sample, often via a metallic transducer, and the echoes returning from buried interfaces are detected via reflectivity changes at the sample surface.

We have used this technology to map an X-ray Bragg mirror consisting of 60 silicon/molybdenum (Si/Mo) layers sputtered onto a monocrystalline Si wafer to investigate post-manufacturing growth homogeneity. The nominal layer period is 6.8 nm and the total stack thickness is 408 nm. The result of a 100x100 pixel scan can be seen in Figure 1.




Fig. 6. Absorption spectrum of atmospheric air (black lines) compared to data compiled from HITRAN database (red lines) for 60 s acquisition time.





Fig. 7. Spectral THz transmission through a 2D electron gas sample in a transient multitesla magnetic field.

#### **ASOPS ENGINE LASER DATASHEET**

Specification*	ASOPS Engine GHz	ASOPS Engine MHz
Repetition Rate	1 GHz	84 MHz**
Typical Repetition Rate Off- set	2 kHz to 20 kHz	10 Hz to 1 kHz
Time Resolution	<60 fs* Over Full 1 ns Window	<100 fs* Over 5 ns Window
Time Delay Window	1 ns	11.9 ns


<sup>\*</sup>Time resolution inherently increases at larger offset frequencies.

# **DIMENSIONS (MM)**



Drawings are for illustrative purposes only, please contact us for complete engineer's drawings

# TL-1000-ASOPS



## SYSTEM CONFIGURATION

The ASOPS Engine consists of two femtosecond lasers (purchased separately), a TL-1000 ASOPS for offset frequency stabilisation, an optical trigger unit. The femtosecond lasers used can be chosen from either the 1 GHz or the 84 MHz range of lasers (purchased separately).

### ADDITIONAL INFORMATION

- Enables two-colour pump-probe experiments
- Available at repetition rates of 84 MHz and 1 GHz
- Enables high-speed data acquisition up to 20 kHz
- Turn-key locking electronics
- Offset frequencies between 4 µHz and 20 kHz

## **CONTACT US**

#### Americas & Asia Pacific

Novanta Headquarters Bedford, USA P +1-781-266-5700

Photonics@Novanta.com

#### Europe, Middle East, Africa

Novanta Europe GmbH Wackersdorf, Germany P +49 9431 7984-0

Milan, Italy P +39-039-793-710

Photonics@Novanta.com

#### China

Novanta Sales & Service Office Shenzhen, China P +86-755-8280-538

Suzhou, China P +86-512-6283-7080

Photonics.China@Novanta.com

#### Japan

Novanta Service & Sales Office Tokyo, Japan P +81-3-5753-2460

Photonics.Japan@Novanta.com

<sup>\*\*</sup> The standard repetition rate is chosen to be 84 MHz. Customer specific repetition rates between 80 MHz and 90 MHz are available upon request.